

Configuration

General

Idle Timeout ON OFF default is 2 minutes
 Ancillary Time Out ON OFF default is 2 minutes
 Password default is 000000

Network Configuration

DHCP Address supplied by network server IP Address
 AUTO Address supplied by operating system IP Netmask
 MANUAL Address is user defined IP Gateway

WAV File Resolution

No WAV file Low Res Medium Res High Res Noise Camera Compatible

Trigger Parameters

Duration Level (SPL)
 SPL A LEQ A On period (sec) Record PreTrigger (sec)
 SPL C LEQ C Off Period (sec) Record PostTrigger (sec)

Percentile Level (SPL)
 SPL A LEQ A Req Ln (%) Record PreTrigger (sec)
 SPL C LEQ C Stat Duration (min) Record PostTrigger (sec)

RMS Detection Level (RMS) Record PostTrigger (sec) Record PreTrigger (sec)

Timer Recurrence Hourly Start Time (mm:ss) Record PreTrigger (mm:ss)
 Daily Start Time (hh:mm:ss) Record PreTrigger (hh:mm:ss)

Remote Switch Record PostTrigger (sec) Record PreTrigger (sec)

Alarm Clock

Recurring Weekly Specify Day
 One Shot Daily Start Time Start Time End time No End

Control
 Logger
 EthernetPort

AUX Output

Duration Level (SPL) On period (sec) Off Period (sec)
 SPL A LEQ A
 SPL C LEQ C

Percentile Level (SPL) Req Ln (%) Stat Duration (min)
 SPL A LEQ A
 SPL C LEQ C

RMS Detection Level (RMS)

Timer Recurrence Hourly Start Time (mm:ss) Record PreTrigger (mm:ss)
 Daily Start Time (hh:mm:ss) Record PreTrigger (hh:mm:ss)

Follow Alarm

UDP Output

UDP Comms Config IP Address Port (default 3002)
 Data Output (Select One) SPL-A SPL-C Leq-A Leq-C
 Output Raw PCM Yes No

Advanced Network Settings

Simple Network Time Protocol Primary Server Secondary Server
 Remote Server Server Username
 FTP User Settings Server Username Password